

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082918\
 Data File : VN050948.D
 Acq On : 29 Aug 2018 14:04
 Operator : MD\SY
 Sample : PB112452ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112452ZHE#04

Quant Time: Aug 30 01:25:40 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	589091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	979178	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	823326	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	287242	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	384856	54.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.54%	
35) Dibromofluoromethane	7.59	113	366254	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
50) Toluene-d8	10.09	98	1302705	44.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.46%	
62) 4-Bromofluorobenzene	12.40	95	360734	37.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.54%	
Target Compounds						
16) Acetone	3.83	43	13844	1.214	ug/l	99
18) Methyl Acetate	4.34	43	14604	2.181	ug/l	96
20) Methylene Chloride	4.55	84	39828	4.983	ug/l	99
43) Isopropyl Acetate	8.17	43	351686	30.816	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082918\
Data File : VN050948.D
Acq On : 29 Aug 2018 14:04
Operator : MD\SY
Sample : PB112452ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112452ZHE#04

Quant Time: Aug 30 01:25:40 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

