

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082918\
 Data File : VN050935.D
 Acq On : 29 Aug 2018 8:40
 Operator : MD\SY
 Sample : VN0829WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0829WBL01

Quant Time: Aug 29 14:42:37 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	605102	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	955236	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	799711	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	256193	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	379259	52.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.12%	
35) Dibromofluoromethane	7.59	113	369649	48.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.28%	
50) Toluene-d8	10.09	98	1323645	46.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.18%	
62) 4-Bromofluorobenzene	12.40	95	339425	35.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	71.90%	


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082918\
Data File : VN050935.D
Acq On : 29 Aug 2018 8:40
Operator : MD\SY
Sample : VN0829WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0829WBL01

Quant Time: Aug 29 14:42:37 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. -0.00 min

Lab File: VN050935.D

Acq: 29 Aug 2018 8:40

Instrument :

MSVOA_N

ClientSampleId :


VN0829WBL01

Tot Ion:168 Resp: 605102

Ion Ratio Lower Upper

168 100

99 48.9 39.6 59.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67

250000

200000

150000



100000

50000

0

7.60 7.70 7.80

Time-->



#33

1,2-Dichloroethane-d4

Concen: 52.063 ug/l

RT: 8.03 min Scan# 2002

Delta R.T. -0.00 min

Lab File: VN050935.D


Acq: 29 Aug 2018 8:40

Tgt Ion: 65 Resp: 379259

Ion Ratio Lower Upper

65 100

67 55.2 0.0 110.4



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO

8.03

150000


100000


50000

0

7.90 8.00 8.10

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2176

Delta R.T. -0.00 min

Lab File: VN050935.D

Acq: 29 Aug 2018 8:40

Instrument :

MSVOA_N

ClientSampleId :

VN0829WBL01


Tot Ion:114 Resp: 955236

Ion Ratio Lower Upper

114 100

63 19.3 0.0 38.6




88 15.4 0.0 30.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 48.641 ug/l

RT: 7.59 min Scan# 1867

Delta R.T. -0.00 min

Lab File: VN050935.D

Acq: 29 Aug 2018 8:40


Tgt Ion:113 Resp: 369649

Ion Ratio Lower Upper

113 100

111 102.1 82.6 123.8



192 23.0 18.0 27.0




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 46.587 ug/l

RT: 10.09 min Scan# 2644

Delta R.T. -0.00 min

Lab File: VN050935.D

Acq: 29 Aug 2018 8:40

Instrument :

MSVOA_N

ClientSampleId :





VN0829WBL01

Tot Ion: 98 Resp: 1323645

Ion Ratio Lower Upper

98 100

100 63.3 51.5 77.3



#62

4-Bromofluorobenzene

Concen: 35.950 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.00 min

Lab File: VN050935.D

Acq: 29 Aug 2018 8:40




Tgt Ion: 95 Resp: 339425


Ion Ratio Lower Upper

95 100

174 93.7 0.0 183.0

176 88.9 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 3053
Delta R.T. -0.00 min
Lab File: VN050935.D
Acq: 29 Aug 2018 8:40

Instrument :
MSVOA_N
ClientSampleId :
VN0829WBL01

Tot Ion:117 Resp: 799711
Ion Ratio Lower Upper
117 100
82 52.3 42.6 64.0
119 32.5 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3656
Delta R.T. 0.00 min
Lab File: VN050935.D
Acq: 29 Aug 2018 8:40

Tgt Ion:152 Resp: 256193
Ion Ratio Lower Upper
152 100
115 56.3 27.8 83.4
150 156.3 0.0 351.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

