

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082919\
 Data File : VN057560.D
 Acq On : 29 Aug 2019 20:02
 Operator : JC/SP
 Sample : K4528-31
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-12-13-S1-(0)

Quant Time: Aug 30 13:01:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 08:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	649849	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1010733	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	862548	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	222636	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	382469	37.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.86%	
35) Dibromofluoromethane	7.58	113	313455	35.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.46%	
50) Toluene-d8	10.08	98	1237059	38.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.16%	
62) 4-Bromofluorobenzene	12.40	95	324071	29.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	59.74%	
Target Compounds						
43) Isopropyl Acetate	8.16	43	83464	6.173	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082919\
Data File : VN057560.D
Acq On : 29 Aug 2019 20:02
Operator : JC/SP
Sample : K4528-31
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-S1-(0)

Quant Time: Aug 30 13:01:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082919W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 08:56:38 2019
Response via : Initial Calibration

