

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082919\
 Data File : VN057570.D
 Acq On : 30 Aug 2019 00:16
 Operator : JC/SP
 Sample : K4529-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-12-13-T2-(0)

Quant Time: Aug 30 14:57:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 08:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	571464	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	893499	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	821988	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	217467	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	350656	39.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	78.04%	
35) Dibromofluoromethane	7.58	113	272735	35.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	70.32%	
50) Toluene-d8	10.08	98	1123633	39.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.28%	
62) 4-Bromofluorobenzene	12.40	95	309158	32.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	64.46%	
Target Compounds						
16) Acetone	3.81	43	45746	12.099	ug/l	93
43) Isopropyl Acetate	8.16	43	100837	8.436	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082919\
Data File : VN057570.D
Acq On : 30 Aug 2019 00:16
Operator : JC/SP
Sample : K4529-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-T2-(0)

Quant Time: Aug 30 14:57:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082919W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 08:56:38 2019
Response via : Initial Calibration

