

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082923\
 Data File : VN079094.D
 Acq On : 29 Aug 2023 15:17
 Operator : JC\MD
 Sample : PB155153ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155153ZHE#03

Quant Time: Aug 29 23:30:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	144048	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	274788	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	276793	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	80254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	117667	52.495	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.980%	
35) Dibromofluoromethane	8.177	113	96094	51.950	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.900%	
50) Toluene-d8	10.571	98	363288	52.894	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.780%	
62) 4-Bromofluorobenzene	12.853	95	104726	46.650	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.300%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	24305	26.279	ug/l	89
20) Methylene Chloride	5.283	84	18731	9.503	ug/l #	82
25) 2-Butanone	7.488	43	8550	5.766	ug/l #	58
43) Isopropyl Acetate	8.694	43	151854	27.597	ug/l #	78
51) 4-Methyl-2-Pentanone	10.453	43	30253	9.270	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082923\
Data File : VN079094.D
Acq On : 29 Aug 2023 15:17
Operator : JC\MD
Sample : PB155153ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155153ZHE#03

Quant Time: Aug 29 23:30:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

