

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082923\
 Data File : VN079096.D
 Acq On : 29 Aug 2023 16:06
 Operator : JC\MD
 Sample : PB155153ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155153ZHE#05

Quant Time: Aug 29 23:30:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	143006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	278267	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	279323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	84784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	116317	52.270	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.540%		
35) Dibromofluoromethane	8.177	113	91774	48.995	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.980%		
50) Toluene-d8	10.571	98	362729	52.153	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 104.300%		
62) 4-Bromofluorobenzene	12.853	95	104262	45.862	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 91.720%		
Target Compounds						
						Qvalue
16) Acetone	4.448	43	24411	26.586	ug/l #	82
20) Methylene Chloride	5.283	84	19113	9.768	ug/l #	80
25) 2-Butanone	7.500	43	8445	5.737	ug/l #	74
43) Isopropyl Acetate	8.694	43	165037	29.618	ug/l #	76
51) 4-Methyl-2-Pentanone	10.453	43	32255	9.760	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082923\
Data File : VN079096.D
Acq On : 29 Aug 2023 16:06
Operator : JC\MD
Sample : PB155153ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155153ZHE#05

Quant Time: Aug 29 23:30:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

