

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082923\
 Data File : VN079101.D
 Acq On : 29 Aug 2023 18:07
 Operator : JC\MD
 Sample : PB155153ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155153ZHE#10

Quant Time: Aug 29 23:31:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	140277	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	278424	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	276294	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	80730	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	116372	53.312	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.620%	
35) Dibromofluoromethane	8.171	113	95235	50.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.620%	
50) Toluene-d8	10.571	98	363267	52.200	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.400%	
62) 4-Bromofluorobenzene	12.853	95	106239	46.706	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.420%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	23348	25.923	ug/l	93
20) Methylene Chloride	5.289	84	16125	8.401	ug/l #	86
43) Isopropyl Acetate	8.694	43	208995	37.486	ug/l #	72
51) 4-Methyl-2-Pentanone	10.453	43	27913	8.441	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082923\
Data File : VN079101.D
Acq On : 29 Aug 2023 18:07
Operator : JC\MD
Sample : PB155153ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155153ZHE#10

Quant Time: Aug 29 23:31:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

