

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082924\
 Data File : VN083552.D
 Acq On : 29 Aug 2024 13:21
 Operator : JC\MD
 Sample : P3633-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DSN003

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/30/2024
 Supervised By : Mahesh Dadoda 08/30/2024

Quant Time: Aug 29 17:54:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	143905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	296000	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	311262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	151181	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112054	54.705	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.420%			
35) Dibromofluoromethane	8.165	113	92127	49.864	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.720%			
50) Toluene-d8	10.565	98	325598	47.244	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 94.480%			
62) 4-Bromofluorobenzene	12.847	95	153385	57.087	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 114.180%			
Target Compounds						
16) Acetone	4.442	43	4708m	5.874	ug/l	
30) Chloroform	7.971	83	8815	2.747	ug/l	92
52) Toluene	10.635	92	1136	0.216	ug/l	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082924\
Data File : VN083552.D
Acq On : 29 Aug 2024 13:21
Operator : JC\MD
Sample : P3633-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DSN003

Manual Integrations
APPROVED

Reviewed By : John Carlone 08/30/2024
Supervised By : Mahesh Dadoda 08/30/2024

Quant Time: Aug 29 17:54:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

