

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN082924\  
 Data File : VN083559.D  
 Acq On : 29 Aug 2024 16:26  
 Operator : JC\MD  
 Sample : PB163049TB  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB163049TB

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 08/30/2024  
 Supervised By : Mahesh Dadoda 08/30/2024

Quant Time: Aug 30 03:29:30 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N080724W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Aug 08 06:30:41 2024  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|--------|--------|----------|
| -----                       |                |      |                     |        |        |          |
| Internal Standards          |                |      |                     |        |        |          |
| 1) Pentafluorobenzene       | 8.224          | 168  | 142618              | 50.000 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.100          | 114  | 293179              | 50.000 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.865         | 117  | 314269              | 50.000 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.794         | 152  | 159791              | 50.000 | ug/l   | 0.00     |
|                             |                |      |                     |        |        |          |
| System Monitoring Compounds |                |      |                     |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.577          | 65   | 112437              | 55.388 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 74 - 125 |      | Recovery = 110.780% |        |        |          |
| 35) Dibromofluoromethane    | 8.165          | 113  | 89770               | 49.056 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = 98.120%  |        |        |          |
| 50) Toluene-d8              | 10.565         | 98   | 327969              | 48.046 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 86 - 113 |      | Recovery = 96.100%  |        |        |          |
| 62) 4-Bromofluorobenzene    | 12.847         | 95   | 158974              | 59.737 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 77 - 121 |      | Recovery = 119.480% |        |        |          |
|                             |                |      |                     |        |        |          |
| Target Compounds            |                |      |                     |        | Qvalue |          |
| 16) Acetone                 | 4.436          | 43   | 22949               | 28.891 | ug/l   | 99       |
| 20) Methylene Chloride      | 5.265          | 84   | 17194               | 9.405  | ug/l # | 83       |
| 43) Isopropyl Acetate       | 8.688          | 43   | 253330m             | 46.926 | ug/l   |          |
| -----                       |                |      |                     |        |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN082924\  
Data File : VN083559.D  
Acq On : 29 Aug 2024 16:26  
Operator : JC\MD  
Sample : PB163049TB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
PB163049TB

Manual Integrations  
APPROVED

Reviewed By : John Carlone 08/30/2024  
Supervised By : Mahesh Dadoda 08/30/2024

Quant Time: Aug 30 03:29:30 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N080724W.M  
Quant Title : SW846 8260  
QLast Update : Thu Aug 08 06:30:41 2024  
Response via : Initial Calibration

