

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082924\
 Data File : VN083548.D
 Acq On : 29 Aug 2024 10:57
 Operator : JC\MD
 Sample : VN0829MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0829MBL01

Quant Time: Aug 30 01:21:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	140833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	285212	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	292820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	141558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110075	54.911	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.820%	
35) Dibromofluoromethane	8.165	113	88661	49.803	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.600%	
50) Toluene-d8	10.565	98	310259	46.721	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 93.440%	
62) 4-Bromofluorobenzene	12.847	95	142733	55.132	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 110.260%	


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082924\
Data File : VN083548.D
Acq On : 29 Aug 2024 10:57
Operator : JC\MD
Sample : VN0829MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Quant Time: Aug 30 01:21:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083548.D
Acq: 29 Aug 2024 10:57

Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01






Tgt Ion:168 Resp: 140833
Ion Ratio Lower Upper
168 100
99 65.3 48.2 72.4



#33
1,2-Dichloroethane-d4
Concen: 54.911 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.005 min
Lab File: VN083548.D
Acq: 29 Aug 2024 10:57

Tgt Ion: 65 Resp: 110075
Ion Ratio Lower Upper
65 100
67 50.9 0.0 104.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN083548.D
Acq: 29 Aug 2024 10:57




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01

Tgt Ion:114 Resp: 285212
Ion Ratio Lower Upper
114 100
63 23.1 0.0 44.6
88 16.3 0.0 31.4



#35
Dibromofluoromethane
Concen: 49.803 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN083548.D
Acq: 29 Aug 2024 10:57

Tgt Ion:113 Resp: 88661
Ion Ratio Lower Upper
113 100
111 102.8 82.4 123.6
192 17.3 14.9 22.3









#50
Toluene-d8
Concen: 46.721 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN083548.D
Acq: 29 Aug 2024 10:57




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion: 98 Resp: 310259
Ion Ratio Lower Upper
98 100
100 65.8 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 55.132 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083548.D
Acq: 29 Aug 2024 10:57

Tgt Ion: 95 Resp: 142733
Ion Ratio Lower Upper
95 100
174 69.8 0.0 159.2
176 66.8 0.0 147.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN083548.D
Acq: 29 Aug 2024 10:57

Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01

Tgt Ion:117 Resp: 292820
Ion Ratio Lower Upper
117 100
82 59.9 47.5 71.3
119 30.5 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN083548.D
Acq: 29 Aug 2024 10:57

Tgt Ion:152 Resp: 141558
Ion Ratio Lower Upper
152 100
115 63.0 30.6 91.6
150 156.7 0.0 348.6

