

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082925\
 Data File : VN087694.D
 Acq On : 29 Aug 2025 12:42
 Operator : JC\MD
 Sample : VN0829MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0829MBL01

Quant Time: Aug 29 22:52:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.206	168	149197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	346944	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	324171	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	144779	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	169537	55.461	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.920%			
35) Dibromofluoromethane	8.153	113	127588	50.935	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.860%			
50) Toluene-d8	10.547	98	461268	49.199	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.400%			
62) 4-Bromofluorobenzene	12.829	95	149763	44.917	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 89.840%			
Target Compounds						Qvalue
5) Bromomethane	2.995	94	1038	0.797	ug/l #	50
10) Methyl Iodide	4.583	142	241	6.944	ug/l #	77
13) Acrolein	4.183	56	885	1.357	ug/l	94
15) Acrylonitrile	5.718	53	1150	0.768	ug/l #	76
16) Acetone	4.430	43	3287	1.975	ug/l	95
18) Methyl Acetate	5.012	43	772	0.222	ug/l #	54
20) Methylene Chloride	5.259	84	497	Below Cal	#	26
21) trans-1,2-Dichloroethene	5.777	96	574	0.246	ug/l #	20
23) Vinyl Acetate	6.600	43	2518	0.329	ug/l #	72
25) 2-Butanone	7.483	43	1472	0.672	ug/l #	50
29) Tetrahydrofuran	7.841	42	891	0.632	ug/l #	58
31) Cyclohexane	8.206	56	5424	1.411	ug/l #	1
36) 1,1-Dichloropropene	8.200	75	15354	3.995	ug/l #	52
38) Carbon Tetrachloride	8.200	117	17753	4.679	ug/l #	16
39) Methylcyclohexane	9.588	83	871	0.206	ug/l #	86
51) 4-Methyl-2-Pentanone	10.435	43	1407	0.284	ug/l #	75
58) 2-Chloroethyl Vinyl ether	10.141	63	622	0.254	ug/l #	49
59) 2-Hexanone	11.029	43	2295	0.734	ug/l	75
68) m/p-Xylenes	12.059	106	1424	0.278	ug/l	80
76) 1,2,3-Trichloropropane	12.829	75	88607	26.098	ug/l #	1
81) trans-1,4-Dichloro-2-b...	12.829	75	88607	69.092	ug/l #	1
85) sec-Butylbenzene	13.600	105	2678	0.218	ug/l	97
87) 1,3-Dichlorobenzene	13.717	146	1191	0.219	ug/l	81
88) 1,4-Dichlorobenzene	13.717	146	1191	0.211	ug/l	82
89) n-Butylbenzene	14.035	91	2109	0.210	ug/l	92
90) Hexachloroethane	14.306	117	422	0.219	ug/l #	22
93) 1,2,4-Trichlorobenzene	15.376	180	1415	0.457	ug/l	79
94) Hexachlorobutadiene	15.476	225	795	0.721	ug/l #	57
95) Naphthalene	15.617	128	5024	0.458	ug/l #	76
96) 1,2,3-Trichlorobenzene	15.811	180	1488	0.492	ug/l #	17


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082925\
Data File : VN087694.D
Acq On : 29 Aug 2025 12:42
Operator : JC\MD
Sample : VN0829MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01

Quant Time: Aug 29 22:52:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. 0.000 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion:168 Resp: 149197
Ion Ratio Lower Upper
168 100
99 65.7 57.2 85.8



#5
Bromomethane
Concen: 0.797 ug/l
RT: 2.995 min Scan# 177
Delta R.T. 0.023 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42


Tgt Ion: 94 Resp: 1038
Ion Ratio Lower Upper
94 100
96 45.8 75.4 113.2#








#10
Methyl Iodide
Concen: 6.944 ug/l
RT: 4.583 min Scan# 446
Delta R.T. 0.006 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01







Tgt Ion:142 Resp: 241
Ion Ratio Lower Upper
142 100
127 60.6 38.0 57.0#
141 0.0 12.0 18.0#



#13
Acrolein
Concen: 1.357 ug/l
RT: 4.183 min Scan# 379
Delta R.T. 0.012 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion: 56 Resp: 885
Ion Ratio Lower Upper
56 100
55 76.7 57.3 85.9









#15
Acrylonitrile
Concen: 0.768 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.012 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion: 53 Resp: 1150
Ion Ratio Lower Upper
53 100
52 65.4 67.6 101.4#
51 19.6 30.1 45.1#



#16
Acetone
Concen: 1.975 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion: 43 Resp: 3287
Ion Ratio Lower Upper
43 100
58 25.8 22.6 33.8









#18
Methyl Acetate
Concen: 0.222 ug/l
RT: 5.012 min Scan# 51
Delta R.T. -0.000 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion: 43 Resp: 772
Ion Ratio Lower Upper
43 100
74 0.0 17.6 26.4#



#20
Methylene Chloride
Concen: Below Cal
RT: 5.259 min Scan# 562
Delta R.T. -0.006 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion: 84 Resp: 497
Ion Ratio Lower Upper
84 100
49 28.1 103.4 155.2#
51 58.2 33.8 50.8#
86 0.7 49.6 74.4#









#21
trans-1,2-Dichloroethene
Concen: 0.246 ug/l
RT: 5.777 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion: 96 Resp: 574
Ion Ratio Lower Upper
96 100
61 47.7 118.0 177.0#
98 0.0 50.0 75.0#



#23
Vinyl Acetate
Concen: 0.329 ug/l
RT: 6.600 min Scan# 790
Delta R.T. 0.012 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion: 43 Resp: 2518
Ion Ratio Lower Upper
43 100
86 0.0 8.2 12.4#









#25
2-Butanone
Concen: 0.672 ug/l
RT: 7.483 min Scan# 94
Delta R.T. 0.018 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion: 43 Resp: 1472
Ion Ratio Lower Upper
43 100
72 0.0 19.9 29.9#



#29
Tetrahydrofuran
Concen: 0.632 ug/l
RT: 7.841 min Scan# 1001
Delta R.T. 0.018 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion: 42 Resp: 891
Ion Ratio Lower Upper
42 100
72 22.1 33.7 50.5#
71 6.6 31.9 47.9#









#31
Cyclohexane
Concen: 1.411 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. -0.030 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion: 56 Resp: 5424
Ion Ratio Lower Upper
56 100
69 134.8 23.7 35.5#
84 157.6 64.7 97.1#



#33
1,2-Dichloroethane-d4
Concen: 55.461 ug/l
RT: 8.559 min Scan# 1123
Delta R.T. -0.000 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion: 65 Resp: 169537
Ion Ratio Lower Upper
65 100
67 52.8 0.0 100.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN087694.D

Acq: 29 Aug 2025 12:42

Instrument :

MSVOA_N

ClientSampleId :

VN0829MBL01





Tgt Ion:114 Resp: 346944

Ion Ratio Lower Upper

114 100

63 25.7 0.0 45.2

88 16.2 0.0 33.4



#35

Dibromofluoromethane

Concen: 50.935 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. -0.000 min

Lab File: VN087694.D

Acq: 29 Aug 2025 12:42




Tgt Ion:113 Resp: 127588


Ion Ratio Lower Upper

113 100

111 105.9 82.8 124.2

192 16.7 12.8 19.2





#36

1,1-Dichloropropene

Concen: 3.995 ug/l

RT: 8.200 min Scan# 1062

Delta R.T. -0.153 min

Lab File: VN087694.D

Acq: 29 Aug 2025 12:42

Instrument :

MSVOA_N

ClientSampleId :

VN0829MBL01





Tgt Ion: 75 Resp: 15354

Ion Ratio Lower Upper

75 100

110 9.3 15.9 47.6#

77 0.0 24.4 36.6#



#38

Carbon Tetrachloride

Concen: 4.679 ug/l

RT: 8.200 min Scan# 1062

Delta R.T. -0.141 min

Lab File: VN087694.D

Acq: 29 Aug 2025 12:42




Tgt Ion: 117 Resp: 17753


Ion Ratio Lower Upper

117 100

119 6.3 79.0 118.6#

121 0.0 25.0 37.6#









#39
Methylcyclohexane
Concen: 0.206 ug/l
RT: 9.588 min Scan# 1297
Delta R.T. 0.006 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion: 83 Resp: 871
Ion Ratio Lower Upper
83 100
55 84.9 64.6 96.8
98 28.5 38.4 57.6#



#50
Toluene-d8
Concen: 49.199 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. -0.000 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion: 98 Resp: 461268
Ion Ratio Lower Upper
98 100
100 63.6 52.8 79.2









#51
4-Methyl-2-Pentanone
Concen: 0.284 ug/l
RT: 10.435 min Scan# 1442
Delta R.T. 0.012 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion: 43 Resp: 1407
Ion Ratio Lower Upper
43 100
58 22.0 29.6 44.4#



#58
2-Chloroethyl Vinyl ether
Concen: 0.254 ug/l
RT: 10.141 min Scan# 1392
Delta R.T. -0.000 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion: 63 Resp: 622
Ion Ratio Lower Upper
63 100
106 0.0 21.1 31.7#









#59
2-Hexanone
Concen: 0.734 ug/l
RT: 11.029 min Scan# 11
Delta R.T. -0.147 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion: 43 Resp: 2295
Ion Ratio Lower Upper
43 100
58 34.2 25.9 77.8



#62
4-Bromofluorobenzene
Concen: 44.917 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.006 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion: 95 Resp: 149763
Ion Ratio Lower Upper
95 100
174 72.7 0.0 138.4
176 66.0 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion:117 Resp: 324171
Ion Ratio Lower Upper
117 100
82 59.7 47.4 71.2
119 32.2 24.3 36.5



#68
m/p-Xylenes
Concen: 0.278 ug/l
RT: 12.059 min Scan# 1718
Delta R.T. 0.012 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion:106 Resp: 1424
Ion Ratio Lower Upper
106 100
91 180.6 169.3 253.9





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 2008

Delta R.T. 0.006 min

Lab File: VN087694.D

Acq: 29 Aug 2025 12:42

Instrument :

MSVOA_N

ClientSampleId :

VN0829MBL01





Tgt Ion:152 Resp: 144779

Ion Ratio Lower Upper

152 100

115 64.1 32.3 96.8

150 161.3 0.0 351.0



#76

1,2,3-Trichloropropane

Concen: 26.098 ug/l

RT: 12.829 min Scan# 1849

Delta R.T. -0.141 min

Lab File: VN087694.D




Acq: 29 Aug 2025 12:42


Tgt Ion: 75 Resp: 88607

Ion Ratio Lower Upper

75 100

77 1.5 89.1 267.4#









#81
 trans-1,4-Dichloro-2-butene
 Concen: 69.092 ug/l
 RT: 12.829 min Scan# 11
 Delta R.T. 0.112 min
 Lab File: VN087694.D
 Acq: 29 Aug 2025 12:42




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0829MBL01


Tgt Ion: 75 Resp: 88607
 Ion Ratio Lower Upper
 75 100
 53 0.1 101.8 152.6#
 89 0.0 35.6 53.4#



#85
 sec-Butylbenzene
 Concen: 0.218 ug/l
 RT: 13.600 min Scan# 1980
 Delta R.T. 0.006 min
 Lab File: VN087694.D
 Acq: 29 Aug 2025 12:42

Tgt Ion: 105 Resp: 2678
 Ion Ratio Lower Upper
 105 100
 134 17.8 9.6 28.6





#87

1,3-Dichlorobenzene

Concen: 0.219 ug/l

RT: 13.717 min Scan# 2000

Delta R.T. 0.006 min

Lab File: VN087694.D

Acq: 29 Aug 2025 12:42

Instrument :

MSVOA_N

ClientSampleId :

VN0829MBL01





Tgt Ion:146 Resp: 1191

Ion Ratio Lower Upper

146 100

111 27.5 22.1 66.5

148 52.6 32.0 96.2



#88

1,4-Dichlorobenzene

Concen: 0.211 ug/l

RT: 13.717 min Scan# 2000

Delta R.T. -0.071 min

Lab File: VN087694.D

Acq: 29 Aug 2025 12:42




Tgt Ion:146 Resp: 1191


Ion Ratio Lower Upper

146 100

111 27.5 21.0 63.0

148 52.6 32.0 96.0





#89

n-Butylbenzene

Concen: 0.210 ug/l

RT: 14.035 min Scan# 2109

Delta R.T. 0.006 min

Lab File: VN087694.D

Acq: 29 Aug 2025 12:42

Instrument :

MSVOA_N

ClientSampleId :

VN0829MBL01





Tgt Ion: 91 Resp: 2109

Ion Ratio Lower Upper

91 100

92 48.2 25.9 77.8

134 30.4 11.7 35.0



#90

Hexachloroethane

Concen: 0.219 ug/l

RT: 14.306 min Scan# 2100

Delta R.T. -0.006 min

Lab File: VN087694.D




Acq: 29 Aug 2025 12:42


Tgt Ion: 117 Resp: 422

Ion Ratio Lower Upper

117 100

201 0.0 29.1 87.4#









#93
1,2,4-Trichlorobenzene
Concen: 0.457 ug/l
RT: 15.376 min Scan# 2115
Delta R.T. 0.012 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42




Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01


Tgt Ion:180 Resp: 1415
Ion Ratio Lower Upper
180 100
182 74.8 47.6 142.9
145 20.8 16.5 49.5



#94
Hexachlorobutadiene
Concen: 0.721 ug/l
RT: 15.476 min Scan# 2299
Delta R.T. -0.000 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion:225 Resp: 795
Ion Ratio Lower Upper
225 100
223 66.7 32.0 96.2
227 0.0 32.1 96.5#









#95
Naphthalene
Concen: 0.458 ug/l
RT: 15.617 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01

Tgt Ion:128 Resp: 5024
Ion Ratio Lower Upper
128 100
127 2.2 10.6 15.8#
129 3.4 8.6 12.8#



#96
1,2,3-Trichlorobenzene
Concen: 0.492 ug/l
RT: 15.811 min Scan# 2356
Delta R.T. -0.000 min
Lab File: VN087694.D
Acq: 29 Aug 2025 12:42

Tgt Ion:180 Resp: 1488
Ion Ratio Lower Upper
180 100
182 0.0 46.8 140.3#
145 8.9 17.3 52.0#

