

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082925\
 Data File : VN087694.D
 Acq On : 29 Aug 2025 12:42
 Operator : JC\MD
 Sample : VN0829MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0829MBL01

Quant Time: Aug 29 22:52:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	149197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	346944	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	324171	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	144779	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	169537	55.461	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.920%
35) Dibromofluoromethane	8.153	113	127588	50.935	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.860%
50) Toluene-d8	10.547	98	461268	49.199	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.400%
62) 4-Bromofluorobenzene	12.829	95	149763	44.917	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082925\
Data File : VN087694.D
Acq On : 29 Aug 2025 12:42
Operator : JC\MD
Sample : VN0829MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0829MBL01

Quant Time: Aug 29 22:52:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

