

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083018\
 Data File : VN050960.D
 Acq On : 30 Aug 2018 14:12
 Operator : MD\SY
 Sample : PB112481ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112481ZHE#04

Quant Time: Aug 31 04:12:56 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	582448	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	977327	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	815314	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	275548	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	392849	56.03	ug/l	0.00
Spiked Amount						
			Recovery	=	112.06%	
35) Dibromofluoromethane	7.59	113	372649	47.93	ug/l	0.00
Spiked Amount						
			Recovery	=	95.86%	
50) Toluene-d8	10.09	98	1304113	44.86	ug/l	0.00
Spiked Amount						
			Recovery	=	89.72%	
62) 4-Bromofluorobenzene	12.40	95	349929	36.23	ug/l	0.00
Spiked Amount						
			Recovery	=	72.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	13589	1.15	ug/l	93
18) Methyl Acetate	4.34	43	13512	2.04	ug/l #	75
20) Methylene Chloride	4.55	84	40331	5.10	ug/l	97
43) Isopropyl Acetate	8.17	43	330923	29.05	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083018\
Data File : VN050960.D
Acq On : 30 Aug 2018 14:12
Operator : MD\SY
Sample : PB112481ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112481ZHE#04

Quant Time: Aug 31 04:12:56 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

