

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083018\
 Data File : VN050961.D
 Acq On : 30 Aug 2018 14:36
 Operator : MD\SY
 Sample : PB112481ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112481ZHE#05

Quant Time: Aug 31 04:18:38 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	548607	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	922968	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	781909	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	258905	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	369201	55.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.80%	
35) Dibromofluoromethane	7.59	113	351227	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
50) Toluene-d8	10.09	98	1233372	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	
62) 4-Bromofluorobenzene	12.40	95	331806	36.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.74%	
Target Compounds						
18) Methyl Acetate	4.34	43	15730	2.52	ug/l	94
20) Methylene Chloride	4.55	84	38911	5.23	ug/l	98
43) Isopropyl Acetate	8.17	43	337699	31.39	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083018\
Data File : VN050961.D
Acq On : 30 Aug 2018 14:36
Operator : MD\SY
Sample : PB112481ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112481ZHE#05

Quant Time: Aug 31 04:18:38 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

