

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083018\
 Data File : VN050966.D
 Acq On : 30 Aug 2018 16:41
 Operator : MD\SY
 Sample : J4275-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-02-082918-A

Quant Time: Aug 31 05:13:09 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	587470	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	983146	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	800430	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	277877	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	392102	55.44	ug/l	0.00
Spiked Amount			Recovery	=	110.88%	
35) Dibromofluoromethane	7.59	113	372898	47.68	ug/l	0.00
Spiked Amount			Recovery	=	95.36%	
50) Toluene-d8	10.09	98	1310089	44.80	ug/l	0.00
Spiked Amount			Recovery	=	89.60%	
62) 4-Bromofluorobenzene	12.40	95	341685	35.16	ug/l	0.00
Spiked Amount			Recovery	=	70.32%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	34331	13.83	ug/l	96
18) Methyl Acetate	4.34	43	18462	2.76	ug/l	91
20) Methylene Chloride	4.55	84	45910	5.76	ug/l	96
25) 2-Butanone	6.85	43	7819	2.74	ug/l	90
30) Chloroform	7.38	83	17852	1.37	ug/l	92
43) Isopropyl Acetate	8.17	43	326961	28.53	ug/l #	90
95) Naphthalene	15.14	128	38231	8.47	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083018\
Data File : VN050966.D
Acq On : 30 Aug 2018 16:41
Operator : MD\SY
Sample : J4275-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-02-082918-A

Quant Time: Aug 31 05:13:09 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

