

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN083019\
 Data File : VN057583.D
 Acq On : 30 Aug 2019 12:37
 Operator : JC/SP
 Sample : K4577-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW-1

Quant Time: Aug 31 00:18:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N083019W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Aug 31 00:00:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.18	128	80477	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	464324	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	357939	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.02	65	197910	31.92	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	106.40%
60) 4-Bromofluorobenzene	12.40	95	117980	21.68	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	72.27%
63) Toluene-d8	10.08	98	552512	33.33	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	111.10%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN083019\
Data File : VN057583.D
Acq On : 30 Aug 2019 12:37
Operator : JC/SP
Sample : K4577-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW-1

Quant Time: Aug 31 00:18:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\624N083019W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Aug 31 00:00:09 2019
Response via : Initial Calibration

