

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074181.D
 Acq On : 30 Aug 2022 22:04
 Operator : JC\MD
 Sample : PB147340TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB147340TB

Quant Time: Aug 31 07:05:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	185369	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	358302	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	370164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	159063	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	139799	48.617	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.240%
35) Dibromofluoromethane	7.951	113	120573	51.714	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.420%
50) Toluene-d8	10.381	98	417978	45.522	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.040%
62) 4-Bromofluorobenzene	12.669	95	163515	58.560	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.120%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	40478	21.066	ug/l	93
20) Methylene Chloride	5.010	84	19479	5.599	ug/l #	79
37) Ethyl Acetate	7.340	43	53398	8.440	ug/l #	88
43) Isopropyl Acetate	8.498	43	203318	23.906	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074181.D
Acq On : 30 Aug 2022 22:04
Operator : JC\MD
Sample : PB147340TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB147340TB

Quant Time: Aug 31 07:05:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

