

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074182.D
 Acq On : 30 Aug 2022 22:29
 Operator : JC\MD
 Sample : PB147340ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB147340ZHE#01

Quant Time: Aug 31 07:05:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	194694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	354611	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	372609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	151477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	137873	45.650	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.300%
35) Dibromofluoromethane	7.951	113	118279	51.258	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.520%
50) Toluene-d8	10.380	98	411383	45.270	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.540%
62) 4-Bromofluorobenzene	12.674	95	155646	56.322	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.640%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	38840	19.246	ug/l	97
37) Ethyl Acetate	7.334	43	52261	8.347	ug/l	98
43) Isopropyl Acetate	8.498	43	222838	26.474	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074182.D
Acq On : 30 Aug 2022 22:29
Operator : JC\MD
Sample : PB147340ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB147340ZHE#01

Quant Time: Aug 31 07:05:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

