

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074183.D
 Acq On : 30 Aug 2022 22:54
 Operator : JC\MD
 Sample : PB147340ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB147340ZHE#02

Quant Time: Aug 31 07:05:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	186192	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	345339	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	363225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	147705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	136410	47.228	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.460%
35) Dibromofluoromethane	7.957	113	116738	51.948	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.900%
50) Toluene-d8	10.380	98	400239	45.226	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.460%
62) 4-Bromofluorobenzene	12.668	95	154038	57.237	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.480%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	41445	21.474	ug/l	94
20) Methylene Chloride	5.016	84	10868	2.086	ug/l #	86
37) Ethyl Acetate	7.333	43	56085	9.198	ug/l	96
43) Isopropyl Acetate	8.492	43	227999	27.814	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074183.D
Acq On : 30 Aug 2022 22:54
Operator : JC\MD
Sample : PB147340ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB147340ZHE#02

Quant Time: Aug 31 07:05:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

