

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074186.D
 Acq On : 31 Aug 2022 00:08
 Operator : JC\MD
 Sample : PB147340ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB147340ZHE#05

Quant Time: Aug 31 07:06:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	180745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	345798	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	359949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	146739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	141565	50.490	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.980%	
35) Dibromofluoromethane	7.957	113	113267	50.337	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.680%	
50) Toluene-d8	10.380	98	403618	45.547	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 91.100%	
62) 4-Bromofluorobenzene	12.669	95	152496	56.588	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.180%	
Target Compounds						
16) Acetone	4.222	43	38739	20.677	ug/l	Qvalue 90
20) Methylene Chloride	5.016	84	17272	4.883	ug/l #	81
37) Ethyl Acetate	7.345	43	48672	7.971	ug/l #	93
43) Isopropyl Acetate	8.492	43	197201	24.025	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074186.D
Acq On : 31 Aug 2022 00:08
Operator : JC\MD
Sample : PB147340ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB147340ZHE#05

Quant Time: Aug 31 07:06:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

