

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074191.D
 Acq On : 31 Aug 2022 02:11
 Operator : JC\MD
 Sample : PB147312TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB147312TB

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/31/2022
 Supervised By : Semsettin Yesilyurt 08/31/2022

Quant Time: Aug 31 07:06:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	172806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	329196	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	338578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	130647	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	130697	48.756	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.520%		
35) Dibromofluoromethane	7.957	113	107739	50.295	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.580%		
50) Toluene-d8	10.381	98	378872	44.911	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	89.820%		
62) 4-Bromofluorobenzene	12.669	95	135744	52.912	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	105.820%		
Target Compounds						
16) Acetone	4.228	43	33381	18.636	ug/l #	77
20) Methylene Chloride	5.028	84	18217m	5.624	ug/l	
37) Ethyl Acetate	7.340	43	47069	8.098	ug/l #	94
43) Isopropyl Acetate	8.498	43	207164	26.512	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074191.D
Acq On : 31 Aug 2022 02:11
Operator : JC\MD
Sample : PB147312TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB147312TB

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/31/2022
Supervised By :Semsettin Yesilyurt 08/31/2022

Quant Time: Aug 31 07:06:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

