

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074192.D
 Acq On : 31 Aug 2022 02:36
 Operator : JC\MD
 Sample : N4379-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMPOSITE-DRUMS

Quant Time: Aug 31 07:07:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	171476	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	322539	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	328652	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	124244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	133061	50.023	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.040%	
35) Dibromofluoromethane	7.951	113	111313	53.036	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.080%	
50) Toluene-d8	10.381	98	373486	45.186	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 90.380%	
62) 4-Bromofluorobenzene	12.669	95	131277	52.227	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.460%	
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	38639	21.738	ug/l #	82
20) Methylene Chloride	5.028	84	17174	5.228	ug/l	91
37) Ethyl Acetate	7.340	43	38356	6.735	ug/l #	95
43) Isopropyl Acetate	8.498	43	163540	21.361	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074192.D
Acq On : 31 Aug 2022 02:36
Operator : JC\MD
Sample : N4379-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMPOSITE-DRUMS

Quant Time: Aug 31 07:07:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

