

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074197.D
 Acq On : 31 Aug 2022 04:39
 Operator : JC\MD
 Sample : N4421-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-PALM-SP-02-GR

Quant Time: Aug 31 07:08:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	193134	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	366561	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	365460	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	152147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	144496	48.230	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.460%
35) Dibromofluoromethane	7.951	113	122563	51.383	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.760%
50) Toluene-d8	10.380	98	419159	44.622	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.240%
62) 4-Bromofluorobenzene	12.668	95	160704	56.256	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.520%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	45551	22.753	ug/l	91
17) Carbon Disulfide	4.457	76	6821	1.058	ug/l #	91
20) Methylene Chloride	5.016	84	17750	4.608	ug/l #	78
37) Ethyl Acetate	7.339	43	42973	6.639	ug/l #	93
43) Isopropyl Acetate	8.498	43	163554	18.797	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074197.D
Acq On : 31 Aug 2022 04:39
Operator : JC\MD
Sample : N4421-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-PALM-SP-02-GR

Quant Time: Aug 31 07:08:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

