

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074198.D
 Acq On : 31 Aug 2022 05:03
 Operator : JC\MD
 Sample : N4421-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-PALM-SP-03

Quant Time: Aug 31 07:08:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	175188	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	319627	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	318864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	143163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	127427	46.890	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.780%
35) Dibromofluoromethane	7.951	113	107501	51.686	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.380%
50) Toluene-d8	10.381	98	360556	44.020	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.040%
62) 4-Bromofluorobenzene	12.669	95	142019	57.016	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.040%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	61897	34.086	ug/l #	83
20) Methylene Chloride	5.016	84	18718	5.732	ug/l #	75
37) Ethyl Acetate	7.257	43	8730	1.547	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074198.D
Acq On : 31 Aug 2022 05:03
Operator : JC\MD
Sample : N4421-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-PALM-SP-03

Quant Time: Aug 31 07:08:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

