

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074199.D
 Acq On : 31 Aug 2022 05:28
 Operator : JC\MD
 Sample : N4421-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-PALM-SP-03-GR

Quant Time: Aug 31 07:12:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	176249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	330680	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	343099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.609	152	156034	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	137431	50.266	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.540%		
35) Dibromofluoromethane	7.951	113	113400	52.700	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.400%		
50) Toluene-d8	10.380	98	382136	45.095	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 90.180%		
62) 4-Bromofluorobenzene	12.668	95	150744	58.496	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 117.000%		
Target Compounds						
						Qvalue
16) Acetone	4.228	43	43477	23.798	ug/l #	83
20) Methylene Chloride	5.016	84	19008	5.807	ug/l	95
43) Isopropyl Acetate	8.498	43	32562	4.148	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074199.D
Acq On : 31 Aug 2022 05:28
Operator : JC\MD
Sample : N4421-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-PALM-SP-03-GR

Quant Time: Aug 31 07:12:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

