

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
 Data File : VN074200.D
 Acq On : 31 Aug 2022 05:52
 Operator : JC\MD
 Sample : N4423-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Aug 31 07:12:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	183222	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	335851	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	355552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	155477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	137028	48.211	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.420%
35) Dibromofluoromethane	7.951	113	113621	51.990	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.980%
50) Toluene-d8	10.380	98	398088	46.254	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.500%
62) 4-Bromofluorobenzene	12.669	95	156051	59.623	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	119.240%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	35322	18.598	ug/l	97
20) Methylene Chloride	5.022	84	19525	5.711	ug/l	88
37) Ethyl Acetate	7.339	43	41023	6.918	ug/l #	96
43) Isopropyl Acetate	8.498	43	173169	21.722	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074200.D
Acq On : 31 Aug 2022 05:52
Operator : JC\MD
Sample : N4423-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Aug 31 07:12:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

