



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083022\
Data File : VN074168.D
Acq On : 30 Aug 2022 11:14
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Title : SW846 8260
Last Update : Wed Aug 31 07:02:40 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1824

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.6	10609	PASS
75	95	30	60	58.2	26144	PASS
95	95	100	100	100.0	44947	PASS
96	95	5	9	7.3	3288	PASS
173	174	0.00	2	0.7	237	PASS
174	95	50	100	72.9	32749	PASS
175	174	5	9	7.6	2488	PASS
176	174	95	101	96.8	31712	PASS
177	176	5	9	6.9	2202	PASS