

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083023\
 Data File : VN079112.D
 Acq On : 30 Aug 2023 12:58
 Operator : JC\MD
 Sample : VN0830WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0830WBL01

Quant Time: Aug 31 02:15:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N081723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 18 05:21:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	35622	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	186407	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	153345	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.582	65	86555	29.992	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.967%
60) 4-Bromofluorobenzene	12.853	95	53184	23.776	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	79.267%
63) Toluene-d8	10.571	98	222225	31.032	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.433%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083023\
Data File : VN079112.D
Acq On : 30 Aug 2023 12:58
Operator : JC\MD
Sample : VN0830WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0830WBL01

Quant Time: Aug 31 02:15:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N081723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 18 05:21:14 2023
Response via : Initial Calibration

