

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083024\
 Data File : VN083570.D
 Acq On : 30 Aug 2024 12:05
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Aug 31 00:36:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:23:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	112236	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	232851	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	249567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115641	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	103980	66.288	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	132.580%#
35) Dibromofluoromethane	8.165	113	83343	59.468	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	118.940%
50) Toluene-d8	10.565	98	298058	54.448	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.900%
62) 4-Bromofluorobenzene	12.847	95	127725	63.176	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	126.360%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083024\
Data File : VN083570.D
Acq On : 30 Aug 2024 12:05
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Aug 31 00:36:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:23:16 2024
Response via : Initial Calibration

