

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN083024\  
 Data File : VN083572.D  
 Acq On : 30 Aug 2024 13:40  
 Operator : JC\MD  
 Sample : VN0830MBL01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0830MBL01

Quant Time: Aug 31 02:36:09 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N083024W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Aug 31 00:48:10 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	146014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	294290	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	305322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	140623	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107430	52.644	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.280%
35) Dibromofluoromethane	8.165	113	87449	49.371	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.740%
50) Toluene-d8	10.565	98	312005	45.292	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.580%
62) 4-Bromofluorobenzene	12.847	95	140677	55.055	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN083024\  
 Data File : VN083572.D  
 Acq On : 30 Aug 2024 13:40  
 Operator : JC\MD  
 Sample : VN0830MBL01  
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA\_N/MEOH  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0830MBL01

Quant Time: Aug 31 02:36:09 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N083024W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Aug 31 00:48:10 2024  
 Response via : Initial Calibration

