

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083024\
 Data File : VN083579.D
 Acq On : 30 Aug 2024 16:45
 Operator : JC\MD
 Sample : P3720-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Aug 31 02:41:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	139613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	279775	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	290034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	131210	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	103642	53.116	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.240%
35) Dibromofluoromethane	8.165	113	84452	50.153	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.300%
50) Toluene-d8	10.565	98	304552	46.474	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.940%
62) 4-Bromofluorobenzene	12.847	95	134163	55.230	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083024\
Data File : VN083579.D
Acq On : 30 Aug 2024 16:45
Operator : JC\MD
Sample : P3720-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Aug 31 02:41:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

