

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083024\
 Data File : VN083590.D
 Acq On : 31 Aug 2024 03:08
 Operator : JC\MD
 Sample : VN0830WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0830WBL02

Quant Time: Aug 31 05:05:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	146241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	303845	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	321557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	152925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117824	57.648	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.300%
35) Dibromofluoromethane	8.165	113	94855	51.868	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.740%
50) Toluene-d8	10.565	98	343240	48.185	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.360%
62) 4-Bromofluorobenzene	12.847	95	156970	59.500	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	119.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083024\
Data File : VN083590.D
Acq On : 31 Aug 2024 03:08
Operator : JC\MD
Sample : VN0830WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0830WBL02

Quant Time: Aug 31 05:05:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

