

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083118\
 Data File : VN050985.D
 Acq On : 31 Aug 2018 14:01
 Operator : MD\SY
 Sample : PB112510ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112510ZHE#03

Quant Time: Aug 31 23:20:48 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	584029	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	932121	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	774298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	228838	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	372229	52.94	ug/l	0.00
Spiked Amount			Recovery	=	105.88%	
35) Dibromofluoromethane	7.59	113	357789	48.25	ug/l	0.00
Spiked Amount			Recovery	=	96.50%	
50) Toluene-d8	10.09	98	1276228	46.03	ug/l	0.00
Spiked Amount			Recovery	=	92.06%	
62) 4-Bromofluorobenzene	12.40	95	319051	34.63	ug/l	0.00
Spiked Amount			Recovery	=	69.26%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
18) Methyl Acetate	4.34	43	10984	1.65	ug/l	93
20) Methylene Chloride	4.56	84	41125	5.19	ug/l	96
43) Isopropyl Acetate	8.17	43	245650	22.61	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083118\
Data File : VN050985.D
Acq On : 31 Aug 2018 14:01
Operator : MD\SY
Sample : PB112510ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112510ZHE#03

Quant Time: Aug 31 23:20:48 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

