

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083119\
 Data File : VN057591.D
 Acq On : 30 Aug 2019 16:14
 Operator : JC/SP
 Sample : VN0831WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampled :
 VN0831WBL01

Quant Time: Aug 31 01:51:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 11:59:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	550622	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	861049	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	716975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	189557	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.02	65	328976	36.50	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	73.00%	
35) Dibromofluoromethane	7.58	113	264385	34.35	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	68.70%	
50) Toluene-d8	10.08	98	1032402	37.92	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	75.84%	
62) 4-Bromofluorobenzene	12.40	95	278003	30.42	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	60.84%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083119\
 Data File : VN057591.D
 Acq On : 30 Aug 2019 16:14
 Operator : JC/SP
 Sample : VN0831WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0831WBL01

Quant Time: Aug 31 01:51:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 11:59:03 2019
 Response via : Initial Calibration

