

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN083119\
 Data File : VN057596.D
 Acq On : 30 Aug 2019 18:47
 Operator : JC/SP
 Sample : K4517-19
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T2-S-A1-(0)

Quant Time: Aug 31 01:12:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 11:59:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	598449	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	935955	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	795428	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	213584	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	353739	36.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.22%	
35) Dibromofluoromethane	7.58	113	281843	33.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.38%	
50) Toluene-d8	10.08	98	1125439	38.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.06%	
62) 4-Bromofluorobenzene	12.40	95	297868	29.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	59.98%	
Target Compounds						
20) Methylene Chloride	4.53	84	15139	2.064	ug/l #	80
43) Isopropyl Acetate	8.16	43	75743	6.049	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN083119\
Data File : VN057596.D
Acq On : 30 Aug 2019 18:47
Operator : JC/SP
Sample : K4517-19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T2-S-A1-(0)

Quant Time: Aug 31 01:12:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082919W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 11:59:03 2019
Response via : Initial Calibration

