

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083121\
 Data File : VN068368.D
 Acq On : 31 Aug 2021 14:49
 Operator : JC/MD
 Sample : PB138804TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138804TB

Quant Time: Sep 01 02:06:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 31 02:20:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	427908	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	702664	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	771550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	317230	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	235615	53.945	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 107.900%			
35) Dibromofluoromethane	8.024	113	220637	49.525	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.060%			
50) Toluene-d8	10.446	98	878835	52.697	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 105.400%			
62) 4-Bromofluorobenzene	12.734	95	303359	55.428	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 110.860%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	41780	23.566	ug/l	95
20) Methylene Chloride	5.098	84	44515	10.740	ug/l	95
43) Isopropyl Acetate	8.563	43	37758	5.013	ug/l #	78
95) Naphthalene	15.509	128	21461	3.910	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083121\
Data File : VN068368.D
Acq On : 31 Aug 2021 14:49
Operator : JC/MD
Sample : PB138804TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138804TB

Quant Time: Sep 01 02:06:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 31 02:20:58 2021
Response via : Initial Calibration

