

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083121\
 Data File : VN068369.D
 Acq On : 31 Aug 2021 15:12
 Operator : JC/MD
 Sample : PB138834TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138834TB

Quant Time: Sep 01 02:06:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 31 02:20:58 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	414177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	677327	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	754525	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	301552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	230713	54.574	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 109.140%		
35) Dibromofluoromethane	8.024	113	212565	49.498	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 99.000%		
50) Toluene-d8	10.443	98	862760	53.668	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 107.340%		
62) 4-Bromofluorobenzene	12.734	95	294336	55.791	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 111.580%		
Target Compounds						Qvalue
16) Acetone	4.296	43	42318	25.339	ug/l	95
20) Methylene Chloride	5.101	84	50429	12.928	ug/l	94
43) Isopropyl Acetate	8.563	43	40122	5.526	ug/l #	86
95) Naphthalene	15.507	128	7643	2.930	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083121\
Data File : VN068369.D
Acq On : 31 Aug 2021 15:12
Operator : JC/MD
Sample : PB138834TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138834TB

Quant Time: Sep 01 02:06:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 31 02:20:58 2021
Response via : Initial Calibration

