

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083121\
 Data File : VN068370.D
 Acq On : 31 Aug 2021 15:34
 Operator : JC/MD
 Sample : M3593-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ETGI-280

Quant Time: Sep 01 02:06:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 31 02:20:58 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	399109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	662639	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	716150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	292413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	228004	55.970	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.940%
35) Dibromofluoromethane	8.027	113	207956	49.498	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.000%
50) Toluene-d8	10.446	98	829123	52.719	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.440%
62) 4-Bromofluorobenzene	12.734	95	279957	54.241	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.480%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	55955	40.208	ug/l	99
20) Methylene Chloride	5.103	84	31651	7.688	ug/l	96
25) 2-Butanone	7.346	43	10671	5.714	ug/l	100
43) Isopropyl Acetate	8.563	43	35771	5.036	ug/l #	85
95) Naphthalene	15.509	128	7728	2.956	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083121\
Data File : VN068370.D
Acq On : 31 Aug 2021 15:34
Operator : JC/MD
Sample : M3593-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ETGI-280

Quant Time: Sep 01 02:06:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 31 02:20:58 2021
Response via : Initial Calibration

