


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083121\
 Data File : VN068363.D
 Acq On : 31 Aug 2021 11:42
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
 Title : SW846 8260
 Last Update : Tue Aug 31 02:20:58 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3988

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	12170	PASS
75	95	30	60	51.0	37720	PASS
95	95	100	100	100.0	73920	PASS
96	95	5	9	7.0	5152	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	90.7	67056	PASS
175	174	5	9	8.4	5603	PASS
176	174	95	101	96.5	64725	PASS
177	176	5	9	6.8	4378	PASS