

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083122\
 Data File : VN074204.D
 Acq On : 31 Aug 2022 10:10
 Operator : JC\MD
 Sample : VN0831MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0831MBL01

Quant Time: Sep 01 00:20:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	180655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	334409	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	309173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	133694	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	135591	48.384	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.760%
35) Dibromofluoromethane	7.951	113	113620	52.213	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.420%
50) Toluene-d8	10.380	98	364489	42.533	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.060%#
62) 4-Bromofluorobenzene	12.669	95	137147	52.626	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083122\
Data File : VN074204.D
Acq On : 31 Aug 2022 10:10
Operator : JC\MD
Sample : VN0831MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0831MBL01

Quant Time: Sep 01 00:20:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

