

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083122\
 Data File : VN074215.D
 Acq On : 31 Aug 2022 15:34
 Operator : JC\MD
 Sample : N4379-21
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EQUIP-BLANK

Quant Time: Sep 01 00:26:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	200873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	360600	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	353920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	155787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	163280	52.400	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.800%	
35) Dibromofluoromethane	7.957	113	131991	56.250	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 112.500%	
50) Toluene-d8	10.380	98	439173	47.525	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.060%	
62) 4-Bromofluorobenzene	12.669	95	168195	59.852	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 119.700%	
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	5692	2.734	ug/l #	61
25) 2-Butanone	7.269	43	4904	1.472	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083122\
Data File : VN074215.D
Acq On : 31 Aug 2022 15:34
Operator : JC\MD
Sample : N4379-21
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EQUIP-BLANK

Quant Time: Sep 01 00:26:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

