

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090119\
 Data File : VN057630.D
 Acq On : 31 Aug 2019 22:05
 Operator : JC/SP
 Sample : K4517-35RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T4-S-E-(2.5)RE

Quant Time: Sep 01 01:03:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 15:54:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	306782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	535788	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	469944	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	140978	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	258448	37.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.58%	
35) Dibromofluoromethane	7.58	113	158819	33.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.88%	
50) Toluene-d8	10.08	98	679818	37.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.68%	
62) 4-Bromofluorobenzene	12.40	95	206570	30.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.48%	
Target Compounds						
20) Methylene Chloride	4.53	84	9210	1.915	ug/l	91
43) Isopropyl Acetate	8.16	43	57025	4.684	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090119\
Data File : VN057630.D
Acq On : 31 Aug 2019 22:05
Operator : JC/SP
Sample : K4517-35RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T4-S-E-(2.5)RE

Quant Time: Sep 01 01:03:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N090119W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 15:54:32 2019
Response via : Initial Calibration

