

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090120\
 Data File : VN063166.D
 Acq On : 1 Sep 2020 19:41
 Operator : JC/MD
 Sample : L3850-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 228

Quant Time: Sep 02 04:57:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	354578	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	706199	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	707078	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	297381	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	283698	48.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.88%	
35) Dibromofluoromethane	7.55	113	224380	49.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.06%	
50) Toluene-d8	10.06	98	905433	51.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.84%	
62) 4-Bromofluorobenzene	12.38	95	327409	53.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.34%	
Target Compounds						
8) Diethyl Ether	3.38	74	3693	1.374	ug/l	70
11) Tert butyl alcohol	4.72	59	48342	67.944	ug/l #	90
16) Acetone	3.78	43	39111	23.736	ug/l	94
18) Methyl Acetate	4.29	43	6328	1.498	ug/l	94
20) Methylene Chloride	4.51	84	19848	3.830	ug/l	95
43) Isopropyl Acetate	8.13	43	48729	4.401	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090120\
Data File : VN063166.D
Acq On : 1 Sep 2020 19:41
Operator : JC/MD
Sample : L3850-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
228

Quant Time: Sep 02 04:57:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

