

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090120\
 Data File : VN063168.D
 Acq On : 1 Sep 2020 20:29
 Operator : JC/MD
 Sample : L3851-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 238

Quant Time: Sep 02 05:04:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	366380	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	717305	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	719865	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	291201	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	294355	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
35) Dibromofluoromethane	7.55	113	224546	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
50) Toluene-d8	10.06	98	910404	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
62) 4-Bromofluorobenzene	12.38	95	333185	53.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.54%	
Target Compounds						
16) Acetone	3.78	43	34309	20.151	ug/l	99
18) Methyl Acetate	4.29	43	7982m	1.828	ug/l	
20) Methylene Chloride	4.51	84	16941	3.077	ug/l	89
43) Isopropyl Acetate	8.13	43	101237	9.002	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090120\
Data File : VN063168.D
Acq On : 1 Sep 2020 20:29
Operator : JC/MD
Sample : L3851-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
238

Quant Time: Sep 02 05:04:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

