

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090121\
 Data File : VN068383.D
 Acq On : 1 Sep 2021 13:30
 Operator : JC/MD
 Sample : M3611-01 100X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NWB-1528

Quant Time: Sep 02 12:13:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 31 02:20:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	448310	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	715910	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	706386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	294926	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	237376	51.88	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.76%			
35) Dibromofluoromethane	8.024	113	223114	49.15	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.30%			
50) Toluene-d8	10.443	98	858225	50.51	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.02%			
62) 4-Bromofluorobenzene	12.734	95	282688	50.70	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 101.40%			
Target Compounds						
16) Acetone	4.288	43	613755	520.68	ug/l	98
95) Naphthalene	15.509	128	33433	4.97	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090121\
Data File : VN068383.D
Acq On : 1 Sep 2021 13:30
Operator : JC/MD
Sample : M3611-01 100X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NWB-1528

Quant Time: Sep 02 12:13:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 31 02:20:58 2021
Response via : Initial Calibration

