

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090121\
 Data File : VN068384.D
 Acq On : 1 Sep 2021 13:52
 Operator : JC/MD
 Sample : M3611-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NWB-1601

Quant Time: Sep 02 12:14:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 31 02:20:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	426551	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	694245	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	701982	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	309135	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	235154	54.01	ug/l	0.00
Spiked Amount 50.000	Range	61 - 141	Recovery	= 108.02%		
35) Dibromofluoromethane	8.026	113	221344	50.29	ug/l	0.00
Spiked Amount 50.000	Range	69 - 133	Recovery	= 100.58%		
50) Toluene-d8	10.446	98	838024	50.86	ug/l	0.00
Spiked Amount 50.000	Range	65 - 126	Recovery	= 101.72%		
62) 4-Bromofluorobenzene	12.733	95	295975	54.73	ug/l	0.00
Spiked Amount 50.000	Range	58 - 135	Recovery	= 109.46%		
Target Compounds						
						Qvalue
16) Acetone	4.299	43	18376	2.23	ug/l	97
69) o-Xylene	12.280	106	10662	1.19	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090121\
Data File : VN068384.D
Acq On : 1 Sep 2021 13:52
Operator : JC/MD
Sample : M3611-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NWB-1601

Quant Time: Sep 02 12:14:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 31 02:20:58 2021
Response via : Initial Calibration

