


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090121\
 Data File : VN068377.D
 Acq On : 1 Sep 2021 8:31
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
 Title : SW846 8260
 Last Update : Thu Sep 02 14:59:33 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3989

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	15410	PASS
75	95	30	60	50.3	45149	PASS
95	95	100	100	100.0	89768	PASS
96	95	5	9	6.1	5483	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	89.2	80069	PASS
175	174	5	9	8.1	6487	PASS
176	174	95	101	96.4	77176	PASS
177	176	5	9	6.6	5064	PASS