

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
 Data File : VN074230.D
 Acq On : 01 Sep 2022 09:55
 Operator : JC\MD
 Sample : VN0901MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0901MBL01

Quant Time: Sep 01 16:13:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	174537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	321355	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	307646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	127378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	145152	53.611	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.220%	
35) Dibromofluoromethane	7.951	113	117382	56.134	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 112.260%	
50) Toluene-d8	10.380	98	384339	46.671	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 93.340%	
62) 4-Bromofluorobenzene	12.674	95	142599	56.941	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.880%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
Data File : VN074230.D
Acq On : 01 Sep 2022 09:55
Operator : JC\MD
Sample : VN0901MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0901MBL01

Quant Time: Sep 01 16:13:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

