

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
 Data File : VN074237.D
 Acq On : 01 Sep 2022 13:03
 Operator : JC\MD
 Sample : N4421-01
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-PALM-SP-01

Quant Time: Sep 01 16:15:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.004	168	205200	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.898	114	386544	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	387265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	209328	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	147702	46.401	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.800%
35) Dibromofluoromethane	7.945	113	122797	48.820	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.640%
50) Toluene-d8	10.380	98	425600	42.965	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.940%#
62) 4-Bromofluorobenzene	12.668	95	198856	66.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	132.020%#
Target Compounds						
					Qvalue	
25) 2-Butanone	7.275	43	40637	11.943	ug/l	93
80) 1,3,5-Trimethylbenzene	12.998	105	9429	0.647	ug/l	94
84) 1,2,4-Trimethylbenzene	13.310	105	18788	1.335	ug/l	94
85) sec-Butylbenzene	13.433	105	8667	0.495	ug/l	93
86) p-Isopropyltoluene	13.551	119	5998	0.441	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
Data File : VN074237.D
Acq On : 01 Sep 2022 13:03
Operator : JC\MD
Sample : N4421-01
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-PALM-SP-01

Quant Time: Sep 01 16:15:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

