

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
 Data File : VN074238.D
 Acq On : 01 Sep 2022 13:27
 Operator : JC\MD
 Sample : N4421-06
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-PALM-SP-02

Quant Time: Sep 01 16:15:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	242411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	464401	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	482726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	212811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	174176	46.319	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.640%
35) Dibromofluoromethane	7.945	113	142577	47.180	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.360%
50) Toluene-d8	10.375	98	527045	44.286	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.580%
62) 4-Bromofluorobenzene	12.669	95	251692	69.545	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	139.100%#
Target Compounds						
					Qvalue	
25) 2-Butanone	7.275	43	45468	11.311	ug/l #	80
80) 1,3,5-Trimethylbenzene	12.998	105	34329	2.316	ug/l	65
84) 1,2,4-Trimethylbenzene	13.304	105	44735	3.127	ug/l	93
85) sec-Butylbenzene	13.439	105	19689	1.107	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
Data File : VN074238.D
Acq On : 01 Sep 2022 13:27
Operator : JC\MD
Sample : N4421-06
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-PALM-SP-02

Quant Time: Sep 01 16:15:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

